



EXTRACTION FORCEPS





Extraction Forceps No. 1
(For upper centrals, canines and roots)

210101SS	Standard Stainless Steel
210101NB	Nano Coating Black
210101NG	Nano Coating Gold



Extraction Forceps No. 2
(For upper laterals)

210102SS	Standard Stainless Steel
210102NB	Nano Coating Black
210102NG	Nano Coating Gold



Extraction Forceps No. 7
(For upper premolars and roots)

210103SS	Standard Stainless Steel
210103NB	Nano Coating Black
210103NG	Nano Coating Gold



Extraction Forceps No. 29
(For upper roots)

210104SS	Standard Stainless Steel
210104NB	Nano Coating Black
210104NG	Nano Coating Gold



Extraction Forceps No. 29S
(For upper roots; small beak)

210105SS	Standard Stainless Steel
210105NB	Nano Coating Black
210105NG	Nano Coating Gold



Extraction Forceps No. 30
(For upper premolars and roots)

210106SS	Standard Stainless Steel
210106NB	Nano Coating Black
210106NG	Nano Coating Gold



Extraction Forceps No. 44N
(For upper roots; narrow beak)

210107SS	Standard Stainless Steel
210107NB	Nano Coating Black
210107NG	Nano Coating Gold



Extraction Forceps No. 107
(For upper canines)

210108SS	Standard Stainless Steel
210108NB	Nano Coating Black
210108NG	Nano Coating Gold





Extraction Forceps No. 113
(For upper roots)

210109SS	Standard Stainless Steel
210109NB	Nano Coating Black
210109NG	Nano Coating Gold



Extraction Forceps No. 136
(For upper premolars and roots)

210110SS	Standard Stainless Steel
210110NB	Nano Coating Black
210110NG	Nano Coating Gold



Extraction Forceps No. 147
(For upper roots)

210111SS	Standard Stainless Steel
210111NB	Nano Coating Black
210111NG	Nano Coating Gold



Extraction Forceps No. 17
(Right Side - For upper molars)

210112SS	Standard Stainless Steel
210112NB	Nano Coating Black
210112NG	Nano Coating Gold



Extraction Forceps No. 18
(Left Side - For upper molars)

210113SS	Standard Stainless Steel
210113NB	Nano Coating Black
210113NG	Nano Coating Gold



Extraction Forceps No. 19
(For upper wisdoms and roots)

210114SS	Standard Stainless Steel
210114NB	Nano Coating Black
210114NG	Nano Coating Gold



Extraction Forceps No. 51
(For upper wisdoms; Bayonet)

210115SS	Standard Stainless Steel
210115NB	Nano Coating Black
210115NG	Nano Coating Gold



Extraction Forceps No. 67
(For upper wisdoms; Bayonet)

210116SS	Standard Stainless Steel
210116NB	Nano Coating Black
210116NG	Nano Coating Gold





Extraction Forceps No. 76
(For upper roots)

210117SS	Standard Stainless Steel
210117NB	Nano Coating Black
210117NG	Nano Coating Gold



Extraction Forceps No. 76N
(For small upper roots; narrow beak)

210118SS	Standard Stainless Steel
210118NB	Nano Coating Black
210118NG	Nano Coating Gold



Extraction Forceps No. 76S
(For upper roots; small beak)

210119SS	Standard Stainless Steel
210119NB	Nano Coating Black
210119NG	Nano Coating Gold



Extraction Forceps No. 89
(Right Side - for upper molars; Upper Cowhorn)

210120SS	Standard Stainless Steel
210120NB	Nano Coating Black
210120NG	Nano Coating Gold



Extraction Forceps No. 90
(Left Side - for upper molars; Upper Cowhorn)

210121SS	Standard Stainless Steel
210121NB	Nano Coating Black
210121NG	Nano Coating Gold



Extraction Forceps No. 94
(Right Side - for upper molars)

210122SS	Standard Stainless Steel
210122NB	Nano Coating Black
210122NG	Nano Coating Gold



Extraction Forceps No. 95
(Left Side - for upper molars)

210123SS	Standard Stainless Steel
210123NB	Nano Coating Black
210123NG	Nano Coating Gold



Extraction Forceps No. 101
(For upper molars and roots)

210124SS	Standard Stainless Steel
210124NB	Nano Coating Black
210124NG	Nano Coating Gold





Extraction Forceps No. 13
(For lower premolars)

210125SS	Standard Stainless Steel
210125NB	Nano Coating Black
210125NG	Nano Coating Gold



Extraction Forceps No. 74
(For lower anteriors and roots)

210126SS	Standard Stainless Steel
210126NB	Nano Coating Black
210126NG	Nano Coating Gold



Extraction Forceps No. 74N
(For small lower anteriors and roots;
narrow beak)

210127SS	Standard Stainless Steel
210127NB	Nano Coating Black
210127NG	Nano Coating Gold



Extraction Forceps No. 75
(For lower premolars)

210128SS	Standard Stainless Steel
210128NB	Nano Coating Black
210128NG	Nano Coating Gold



Extraction Forceps No. 137
(For lower centrals and roots)

210129SS	Standard Stainless Steel
210129NB	Nano Coating Black
210129NG	Nano Coating Gold



Extraction Forceps No. 22
(For lower molars; Hawk's bill)

210130SS	Standard Stainless Steel
210130NB	Nano Coating Black
210130NG	Nano Coating Gold



Extraction Forceps No. 73
(For lower molars; Hawk's bill)

210131SS	Standard Stainless Steel
210131NB	Nano Coating Black
210131NG	Nano Coating Gold



Extraction Forceps No. 73S
(For lower molars; Hawk's bill; Small
beak)

210132SS	Standard Stainless Steel
210132NB	Nano Coating Black
210132NG	Nano Coating Gold





Extraction Forceps No. 79
(For lower wisdoms)

210133SS	Standard Stainless Steel
210133NB	Nano Coating Black
210133NG	Nano Coating Gold



Extraction Forceps No. 86
(For lower molars; Cowhorn)

210134SS	Standard Stainless Steel
210134NB	Nano Coating Black
210134NG	Nano Coating Gold



Extraction Forceps No. 87
(For lower molars; Cowhorn)

210135SS	Standard Stainless Steel
210135NB	Nano Coating Black
210135NG	Nano Coating Gold



Extraction Forceps No. 37
(Children's upper incisors and canines)

210136SS	Standard Stainless Steel
210136NB	Nano Coating Black
210136NG	Nano Coating Gold



Extraction Forceps No. 138
(Children's upper anteriors, premolars and roots)

210137SS	Standard Stainless Steel
210137NB	Nano Coating Black
210137NG	Nano Coating Gold



Extraction Forceps No. 163
(Children's upper anteriors)

210138SS	Standard Stainless Steel
210138NB	Nano Coating Black
210138NG	Nano Coating Gold



Extraction Forceps No. 39
(Children's upper molars)

210139SS	Standard Stainless Steel
210139NB	Nano Coating Black
210139NG	Nano Coating Gold



Extraction Forceps No. 157
(Children's upper molars)

210140SS	Standard Stainless Steel
210140NB	Nano Coating Black
210140NG	Nano Coating Gold





Extraction Forceps No. 158
(For children's upper molar)

210141SS	Standard Stainless Steel
210141NB	Nano Coating Black
210141NG	Nano Coating Gold



Extraction Forceps No. 159
(For children's upper premolars)

210142SS	Standard Stainless Steel
210142NB	Nano Coating Black
210142NG	Nano Coating Gold



Extraction Forceps No. 123
(For children's lower anteriors, premolars and roots)

210143SS	Standard Stainless Steel
210143NB	Nano Coating Black
210143NG	Nano Coating Gold



Extraction Forceps No. 162
(For children's lower anteriors and roots)

210144SS	Standard Stainless Steel
210144NB	Nano Coating Black
210144NG	Nano Coating Gold



Extraction Forceps No. 160
(For children's lower molars; Hawk's bill)

210145SS	Standard Stainless Steel
210145NB	Nano Coating Black
210145NG	Nano Coating Gold



Extraction Forceps No. 161
(For children's lower molars; Hawk's bill)

210146SS	Standard Stainless Steel
210146NB	Nano Coating Black
210146NG	Nano Coating Gold



Extraction Forceps No. 33

210147SS	Standard Stainless Steel
210147NB	Nano Coating Black
210147NG	Nano Coating Gold





Cleaning: Clean and remove all debris immediately after use. Use an enzymatic cleaning solution in an ultrasonic cleaner to clean instruments. Your ultrasonic solution should be changed daily. Hinged instruments should be cleaned and sterilized in the open position. Discoloration and oxidation may occur if instruments are improperly cleaned. To increase the life of your instruments, clean them well. Use detergents and disinfectants suitable for use with medical and/or dental instruments.

Lubrication: To increase the life of your instruments, routine lubrication is recommended. Prior to sterilization, lubricate instruments. This should be done on daily basis or at least weekly. Use only lubricants designed for precision hinged instruments. If using a dry heat sterilizer, be sure to use a lubricant that is compatible with the operating temperatures of your dry heat unit.

Sharpening: Regular sharpening to your cutters will increase its service life, and its ability to cut more effectively. Frequency of sharpening depends on the frequency of use, as well as, the types and diameters of wires on which the cutter is being used. Check cutters frequently to assess if sharpening is needed.

Sterilising: Instruments should be in the open position when sterilised. The most efficient method is steam autoclave. Anaqa Range of instruments is actually designed for use with steam autoclave as well as Ethylene Oxide (EtO) sterilisation. Steam autoclaves will not dull cutters. Other types of sterilising methods include chemical vapour or dry heat. Using cold sterilants is not recommended. If used, they can chemically attack your instruments.

Corrosion: Anaqa range of instruments is very resistant to corrosion. You can prevent corrosion problems by properly following all manufacturer's recommendations. Discoloration may appear that may not be corrosion. Organic materials that remain can give the appearance of corrosion and be mistaken for rust. An Orange/Brown stain is usually a phosphate layer caused by water sources, cleaning detergents, sterilisation solutions. Black stain is usually acid reaction caused by detergents. Dark brown stain usually caused by dried blood residues. Bluish / black stain usually is plating caused by autoclave cleaning material. Never mix a corroded instrument with new instruments as oxidation may be spread to new instruments. It is recommended that corroded instrument be discarded or replaced.

Maintenance and Care: It is extremely important to follow proper handling, care and sterilization methods to ensure proper instruments function and improve the instrument life. Avoid tap water during the cleaning and sterilization of instruments. Tap water contains high levels of minerals that can cause instrument discoloration and corrosion. Tap water can also neutralize rust inhibiting and lubricating solution. Always use purified, distilled water in ultrasonic cleaning and autoclave sterilization units. It is also recommended to use a no-rinse ultrasonic solution that contains both a rust inhibitor and a lubricant prior to sterilization and follow the manufacturer's exact solution measurements.

Do not leave instruments in the ultrasonic solution after cycle is complete, remove immediately and dry completely. Autoclaves contain high levels of moisture and can be damaging if not functioning properly. Instrument handling is also important. Make sure instruments are dry before sterilization, keep jaws open, and do not overload trays. When cycle is complete, remove instruments immediately after cooling. Always check that joints and tips are moisture free or corrosion can occur. Lubricate with silicone-based lubricant before storage.

Corrosion Prevention: Corrosion can be an issue with any stainless steel instrument. Stainless steel requires Oxygen to form a protective chromium oxide surface layer, which protects against corrosion. Water droplets, organic residue and bonding agents left on the instruments prevent oxygen contact with the surface of instrument and protective layer of chromium oxide will not form. It is recommended

- Instruments should be thoroughly cleaned of all residual matter prior to sterilization.
- All instruments should be sterilized in their open position.
- All instruments should be thoroughly air or towel dried taking special care in the joint areas and crevices.
- Instruments that have corrosion should never be sterilised with non-corroded instruments as the iron oxide on the corroded instrument can transfer to the non-corroded instruments permanently.



5 YEARS	Extraction Forceps
5 YEARS	Orthodontic Pliers & Cutters
1 YEAR	All Periodontic, Hygiene and Plastic Instruments

ISFY guarantees against defects in materials and manufacturing. If your Anaqa instruments should fail or are damaged due to a manufacturing or workmanship defect, we will repair or replace the product without quibble or charge once the faulty product is returned with proof of purchase.

Please note that this guarantee does not cover damage caused by accident, improper care, misuse, alterations to the Anaqa extraction instruments or user negligence.

Anaqa extraction instruments should be disinfected following the processing guidelines provided. This guarantee does not cover damages which arise because these guidelines have not been followed.

ISFY reserves the right to decide whether an instrument is to be repaired or replaced. Where instruments require sharpening, this is considered routine maintenance with normal use and is not covered by any guarantee.

WARRANTY INFORMATION

For more details, please visit www.isfy.co.uk or contact us via email on contact@isfy.co.uk.



ISFY

QUALITY THROUGH INNOVATION

ISFY manufactures high quality, durable single and reusable surgical instruments. We also manufacture equipment to clients' exacting specifications.

We are serious about standards, and for that reason we have sourced the very best materials, machinery and processes from across the globe. For example, we use steel from Japan and tungsten carbide tips from the UK, for our premium Anaqa range.

We also invest in the very best people – from the highly skilled and experienced teams that manufacture our equipment, to the management team of the company, who between them have a wealth of experience within the industry.

We will never compromise on quality, which is why we offer market-leading warranties on our products – up to FIVE years on our Anaqa range. Service is at the heart of everything we do. We pride ourselves on our flexibility and personal touch.

We are constantly looking to improve our processes and explore the latest technologies, materials and techniques to generate what we call “quality through innovation”.

ISFY Ltd
I-Centre, Oakham Business Park
Hamilton Way
Mansfield
Nottinghamshire
NG18 5BR
UK
Tel: +44 (0)115 786 7111
Email: contact@isfy.co.uk
Web: www.isfy.co.uk